



<p>Change Settings</p> <ul style="list-style-type: none"> Adult/Pediatric/Neonatal Target Inflation Pressure Temperature Mode Pulse Tone Volume 	<p>Set Alarm Limits (repeat)</p> <hr/> <p>Set NIBP Interval </p> <hr/> <p>Review Patient Data (after at least one measurement cycle)</p>	
<p>Neonatal Term birth through 28 days, or up to 44 gestational weeks</p> <p>Pediatric 29 days through 12 years</p> <p>Adult 13 years and older</p>	<p>Vital Signs Monitor 300 Series</p>	
<p>WelchAllyn</p>	<p>(800) 289-2500 (503) 530-7500 www.monitoring.welchallyn.com</p>	<p>P/N 810-1710-01 Rev. A 10/2005 ©2005 Made in USA Beaverton, OR USA</p>